

TECHNICAL DATA

Fluke 700HTH-1 Hydraulic Test Hose



Key features

- Hose and fittings connect directly to the Fluke 700HTP-1 test pressure port
- Includes plugs and caps to eliminate fluid loss in hose, eliminates the need to re-prime the hose
- Self-sealing fittings with easy finger tight connections
- Low volumetric expansion and a negligible pressure drop at rated pressures

Product overview: Fluke 700HTH-1 Hydraulic Test Hose

The Fluke 700HTH-1 hydraulic test hose is a 10,000 psi, 690 bar working pressure test hose that features stainless steel and nylon construction for use in connecting a Fluke-700HTP-1 or other hydraulic test pump to a unit under test.

The Fluke 700HTH-1 allows connection to both a Fluke pressure module and a calibration unit under test from a Fluke 700HTP-1 hydraulic test pump simultaneously. 700HTH-1 is compatible with water and non-corrosive oil.

Specifications: Fluke 700HTH-1 Hydraulic Test Hose

Mechanical and General Specifications	
Minimum bend radius	60 mm / 2.36 in.
Temperature	- 40°C to +120°C, (-40°F to +250°F)

Material	Nylon, stainless steel, zinc, plated steel, and nitrile
Dimensions	.84 m (33 in) long without adapter fittings
Weight	412.8 g (.910 oz)
Warranty	One year

Ordering information



Fluke 700HTH-1

Hydraulic Test Hose

-
- Hose with M16X2.0 connectors
 - 2 1/4" male NPT to M16X2.0 connectors
 - 1/4 in NPT female to 1/4 in BSP male adapter
 - Instruction sheet
-

Fluke. *Keeping your world up and running.®*

Fluke Corporation

PO Box 9090, Everett, WA 98206 U.S.A.

Fluke Electronics (Malaysia) Sdn Bhd

Suite D-3A-4, Block D
SouthGage Commercial Centre, No2
Jalan Dua Off Jln Chan Sow Lin
55200 KL
Malaysia

Email: info.asean@fluke.com

www.fluke.com/my

©2025 Fluke Corporation. Specifications subject to
change without notice.

12/2025

For more information call:

In the U.S.A. (800) 443-5853

In Europe/M-East/Africa +31 (0)40 267 5100

In Canada (800)-36-FLUKE

From other countries +1 (425) 446-5500

**Modification of this document is not permitted
without written permission from Fluke
Corporation.**